

Notification of completion of the development of new type Sound Calibrator as well as the start of the sales



In early spring of 2007, our development department has completed the development of new type Sound Calibrator 「TYPE-2127」, complying with JIS C1515(2004) Class 1 and IEC60942 (2003) Class1. Also, we have simultaneously tooled up production of TYPE-2127.

Then we are to start the sale of the Sound Calibrator very recently.

In parallel with the start of 「TYPE-2127」 sale, we have to inform the completion of the sale of 「TYPE-2126」.

- Pursuant Standards : JIS C 1515(2004) Class1, IEC 60942 : 2003 Class1
- Coverage microphone : 1-inch and 1/2-inch measurement microphones
- Stable and high-precision calibration realized by feedback control with temperature compensation (patent pend.).

Specification	
Pursuant Standards	IEC 60942 : 2003 Class1 JIS C1515 : 2004 Class1
Coverage microphone	Measurement Microphone 1-inch and 1/2-inch (Included 1/2-inch microphone Adapter)
Nominal Sound pressure level (Tolerance)	94dB(±0.3dB)
Nominal Frequency (Tolerance)	1000Hz(±0.1%)
Size (mm)	Approx.110(W)×75.5(H)×46(D)
Weight	Approx.150g
Battery Type	Two Alkaline dry batteries (LR6)
Included Accessories	1/2-inch microphone Adapter Alkaline dry batteries (LR6) Case Instruction manual
Option Accessories	AD-2124-4 : 1/4-inch microphone Adapter